

Xiantao Taichen Protective Products Co., Ltd. CE Technical Document Doc. No.: TC/CE 02 V/N: C/1 1. EC Declaration of Conformity We, the manufacturer, herewith declare that the products Single-use Medical Face Mask meet the provisions of MDR 2017/745 which apply to them. The medical device has been assigned to class I according to Rule 1, Annex VIII, Chapter III of EU Medical Device Regulation (EU) 2017/745. It bears the mark following the procedure relating to the EC Declaration of Conformity set out in Annex II of MDR 2017/745. It is confirmed that a sample of the product has been tested and found in conformity with below Test Standard: EN14683: 2019 Medical face masks - Requirements and test methods Classifications: Type IIR The above mentioned declaration of conformity is exclusively under the responsibility of Company: Xiantao Taichen Protective Products Co., LTD Address: No.1 Zhongling Industry Park, Pengchang Town, Xiantao City, Hubei province, China

Käufer erhalten das ungeschwärzte Zertifikat



Vertrauen verpflichtet





中国认可 国际互认 检测 **TESTING CNAS L0599**

age 1 of 3

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description Disposable protective face mask

SGS Internal Ref No. WHIN2005001155SC

Style No. TC0604 Sample Color Blue

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date May 27, 2020

Testing Period May 28, 2020 - Jun 17, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Panjing L. Helen Avan

Dongjing Liu / Hailian Xuan (Authorized Signatory)

f (86-21) 64958763

Member of the SGS Group (SGS SA)

Käufer erhalten das ungeschwärzte Zertifikat



Vertrauen verpflichtet



SL52025258880901TX Date:June 17.2020 Page 2 of 3

Test Result

Medical Face Masks-Requirements and Test Methods

(EN 14683:2019+AC:2019)

Clause 5.2.2 Bacterial filtration efficiency (BFE)

(EN 14683:2019 Annex B)

(BFE), % 99.9 99.9 99.9 99.9

Remark: Performance Requirement: Type I≥95%, Type II≥98%, Type IIR ≥98%

Clause 5.2.3 Breathability

(EN 14683 :2019+AC:2019 Annex C, Flow rate 8 l/min)

	1#	2#	3#	4#	5#
Differential pressure △P (Pa/cm²)	52	49	51	53	51

Remark: Performance Requirement: Type I<40 Pa/cm², Type II<40 Pa/cm², Type IIR<60 Pa/cm²

<u>Clause 5.2.4 Splash Resistance</u> (ISO 22609 :2004, Pressure 16.0 kPa)

1#	2#	3#	4#	5#	6#	7#	8#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
9#	10#	11#	12#	13#	14#	15#	16#
Pass	Pass	Pass	Fail	Pass	Pass	Pass	Pass
17#	18#	19#	20#	21#	22#	23#	24#
Pass	Pass	Pass	Pass	Pass	Pass	Fail	Pass
25#	26#	27#	28#	29#	30#	31#	32#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
umber of P	ass:		30		•	•	
Overall result:		Acceptable					

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR: ≥16.0kPa
- 2) Distance of the medical face mask target area surface to the tip of cannula is 300±10mm.
- 3) Condition and Test temperature (21±5)° C, relative humidity (85±10)%
- 4) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results



中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

Member of the SGS Group (SGS SA)




Vertrauen verpflichtet

Date:June 17,2020 Page 3 of 3 SL52025258880901TX

Clause 5.2.5 Microbial Cleanliness (EN 14683:2019+AC:2019 Annex D and EN ISO 11737-1:2018)

Remark: Performance Requirement: Type I≤30 CFU/g, Type II≤30 CFU/g, Type IIR≤30 CFU/g

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

Member of the SGS Group (SGS SA)